

Urgent Field Safety Notice

VC-20-03.A.OUS

July 2020

Dimension Vista® System

High Sensitivity Troponin I (TNIH) Flex® reagent cartridge Negative Bias with Patient Samples

Our records indicate that your facility may have received the following product:

Table 1. Dimension Vista® affected product

Assay	Test Code	Catalog Number	Siemens Material Number (SMN)	Lot Number	Date of First Distribution	Expiration Date
High Sensitivity Troponin I	TNIH	K6427	10471067	20008BB	2020-02-03	2020-11-03
				20035BC	2020-03-20	2020-11-30
				20135BB	2020-06-02	2021-03-10

Reason for Notice

The purpose of this communication is to inform you of a negative bias with patient samples for Dimension Vista® High Sensitivity Troponin I (TNIH) Flex® reagent cartridge lots listed in Table 1.

Siemens has observed a negative bias across the Analytical Measurement Range of the TNIH assay. The average bias observed for patient samples using the Dimension Vista TNIH lots when compared with a control TNIH lot was -20% for lot 20008BB and -23% for lot 20135BB as shown in Figures 1 and 2. A maximum bias of -26% in patient samples around the 99th percentile with the lots listed in Table 1 was observed. Quality Control will not detect the negative patient bias. TNIH lot 20008BB and lot 20035BC contain the same reagents therefore, performance of the two lots is equivalent.

Siemens is working to restore assay performance, until such time, subsequent lots may contain an alert card in the carton containing lot specific correlation factors.

Figures 1. TNIH % Bias





Figure 2. TNIH Recovery



Risk to Health

A rising or falling pattern in a troponin series for a patient would remain apparent to the clinician when all samples for a patient are tested within the same reagent lot, even if affected by this issue. There is negligible risk to health for this scenario.

The risk for clinical impact is limited to the unlikely scenario where a patient's troponin result when using the affected reagent should have risen above the 99th percentile but did not due to a negative bias when switching from a previous unaffected lot to an affected lot during serial testing. In this case, the potential exists to delay diagnosis of an acute myocardial infarction (AMI) if additional troponin testing was not utilized and other mitigations did not reveal the issue. If present at a higher magnitude than the bias, a rising pattern would be observed in any subsequent samples tested using affected reagent. Initial treatment for suspected AMI would occur, along with consideration of serial troponin testing, electrocardiogram (ECG), clinical history, symptomology and risk factors, including further testing in the setting of an unstable angina diagnosis.

Since troponin test results are used for immediate diagnostic support, Siemens is not recommending a review of previously generated results.

Actions to be Taken by the Customer:

- Calibrate/Recalibrate Dimension Vista TNIH lots listed in Table 1 using any in date lot of Dimension Vista LOCI TNIH Calibrator (SMN 10719482/ Catalogue No. KC627).
- Enter lot specific correlation factors **C0** and **C1** in the Method Configuration screen per Flex reagent Lot listed below in Table 2. After applying correlation factors to the specified lot, product claims listed in the Instructions For Use are met.

Table 2. Correlation Factors for Dimension Vista TNIH lots

TNIH Lot	C0	C1
20008BB	-1.2555	1.2620
20035BC	-1.2555	1.2620
20135BB	0.1387	1.3186

- For instructions on entering correlation factors please refer to the Dimension Vista System Operator's Guide Rev A 2017-03 or online iGuide (Advanced Functions Section); Method Configuration, Section 9.
- Process Quality Control (QC) after recalibration and entry of lot specific correlation factors.

- The correlation factors may create an upward QC shift. Please follow your current laboratory process for adjusting QC ranges. Siemens has provided some examples of adjusted QC values after application of correlation factors. These examples of adjusted QC values for a representative set of QC material are shown in Table 3a and 3b. In order to calculate your own QC bottle values you may apply the correlation factors directly to the QC Insert Sheet Target Value when using the equation below. If the QC Insert Sheet Target Value is not indicated you may use your laboratory's current Target Value.

$$\text{Revised TNIH QC Value} = (\text{TNIH Insert Sheet Target Value} \times C_1) + C_0$$

- Correlation Factors will remain in the TNIH Method configuration until removed manually or revised by the user. New lots of TNIH reagent will require entering of new correlation factors or removal of correlation factors and subsequent adjustment of QC.
- A patient who had serial testing begin before correlation factor implementation may need to be retested after implementation in order to accurately compare results over time. For example, labs many choose to retest one or more previous samples or draw an additional sample.
- Please review this letter with your Medical Director.
- If you received any complaints of illness or adverse events associated with the products listed in Table 1, immediately contact your local Siemens Remote Services Center or your local Siemens Technical Support Representative.
- Complete and return the Field Correction Effectiveness Check Form attached to this letter within 30 days.
- Please retain this letter with your laboratory records and forward this letter to those who may have received this product.

Table 3.a. QC Bottle Values After Application of Correlation Factors for Siemens TNIH Lots 20008BB and 20035BC

BioRad Quality Control Lot	BioRad Insert Sheet Target Value (pg/mL)	Siemens Recovery <u>Without</u> Application of Correlation Factors (pg/mL)	Siemens Recovery <u>With</u> Application of Correlation Factors (pg/mL)	Low Limit -20% Using Correlation Factors (pg/mL)	High Limit +20% Using Correlation factors (pg/mL)
Liquichek™ Cardiac Markers Plus LT 23681	Not assigned	213	268	214	322
Liquichek™ Cardiac Markers Plus LT 23682	Not assigned	4152	5239	4191	6287
Liquichek™ Cardiac Troponin Control 56362	Not assigned	4247	5358	4286	6430
Liquichek™ Cardiac Troponin Control 56363	Not assigned	12276	15491	12393	18589
Liquichek™ Cardiac Markers Plus LT 99562	4872	*	6147	4918	7376
Liquichek™ Cardiac Markers Plus LT 99565	43	*	53	42	64

Table 3.b. QC Bottle Values after application of Correlation Factors for Siemens TNIH Lot 20135BB

BioRad Quality Control Lot	BioRad Insert Sheet Target Value (pg/mL)	Siemens Recovery <u>Without</u> Application of Correlation Factors (pg/mL)	Siemens Recovery <u>With</u> Application of Correlation Factors (pg/mL)	Low Limit -20% Using Correlation Factors (pg/mL)	High Limit +20% Using Correlation Factors (pg/mL)
Liquichek™ Cardiac Markers Plus LT 23681	Not assigned	213	281	225	337
Liquichek™ Cardiac Markers Plus LT 23682	Not assigned	4152	5475	4380	6570
Liquichek™ Cardiac Troponin Control 56362	Not assigned	4247	5600	4480	6720
Liquichek™ Cardiac Troponin Control 56363	Not assigned	12276	16187	12950	19424
Liquichek™ Cardiac Markers Plus LT 99562	4872	*	6424	5139	7709
Liquichek™ Cardiac Markers Plus LT 99565	43	*	57	46	68

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Remote Services Center or your local Siemens Technical Support representative.

Dimension Vista is a trademark of Siemens Healthcare Diagnostics.

FIELD CORRECTION EFFECTIVENESS CHECK

Dimension Vista® High Sensitivity Troponin I (TNIH) Flex® reagent cartridge
Negative Bias with Patient Samples

This response form is to confirm receipt of the enclosed Urgent Field Safety Notice, VC-20-03.A.OUS, dated July 2020 regarding negative bias with patient samples.

Return this completed form to Siemens Healthcare Diagnostics as per the instructions provided at the bottom of this page.

- 1. I have read and understood the Urgent Field Safety Notice instructions provided in this letter. Yes No

- 2. I entered the lot specific correlation factors C0 and C1 in the Dimension Vista System Method Configuration listed in Table 2 and adjusted TNIH QC accordingly. Yes No

Name of person completing questionnaire: _____

Title: _____

Institution: _____ Instrument Serial Number: _____

Street: _____

City: _____ State: _____

Phone: _____ Country: _____

Customer Sold To #: _____ Customer Ship To #: _____

Please send a scanned copy of the completed form via email to your local Siemens Healthineers Technical Support Representative or fax this completed form to your local Siemens Healthineers Technical Support Representative.

If you have any questions, contact your local Siemens Healthineers Technical Support Representative.